

**Annex Attached to Cabinet Resolution No. (83) of 2020 Regarding the UAE System
for Personal Health Protection Products**

No.	Product	Technical Requirements	Approved Standard Specification
1	أقنعة الوجه الطبية Medical Face Masks	<ul style="list-style-type: none"> – Bacterial filtration efficiency (BFE) – Breathability (Differential pressure) – Splash resistance – Microbial cleanliness (Bioburden) – Materials and Construction 	UAE.S EN 14683 (UAE.S ASTM F2100)
2	أقنعة الوجه ذات الفلتر المنقي Half Filtered Face Mask	<ul style="list-style-type: none"> – Total inward leakage – Penetration of filter material – Breathing resistance – Particle Filtration – Compatibility with skin 	UAE.S EN 149
3	أقنعة الوجه غير الطبية (المصنعة من النسيج) Non-Medical "community" Face Mask (made from textile)	<ul style="list-style-type: none"> – pH – Formaldehyde – Heavy metals (Cadmium, Copper, Lead, Nickel) 	UAE.S 1956
4	القفازات الواقية Protective Gloves	<ul style="list-style-type: none"> – Bacteria, Fungi penetration Testing – Protection against Virus 	UAE.S ISO 374-2

No.	Product	Technical Requirements	Approved Standard Specification
5	القفازات الطبية Medical gloves	<ul style="list-style-type: none"> – freedom from holes – Shelf Life and Resistance to degradation – Pack Integrity (for sterile gloves) 	EN 455-Part 1 EN 455-Part 4
6	الملابس الواقية Protective clothing	<ul style="list-style-type: none"> – Resistance to Penetration by blood/body fluids. – Resistance to penetration by viruses. – Resistance to penetration by bacteria. – Resistance to penetration by biologically contaminated aerosols. – Resistance to penetration by contaminated dust. 	EN 14126
7	واقيات العين الشخصية Personal eye-protection (Face shield and Goggles)	<ul style="list-style-type: none"> – Protection against droplets and splashes of liquids – General construction – Materials 	EN 166
8	الأغطية والعبايات الجراحية (Surgical drapes and gowns)	<ul style="list-style-type: none"> – Resistance to microbial penetration – cleanliness microbial/bioburden – Liquid penetration – Tensile strength 	EN 13795 - part 1

No.	Product	Technical Requirements	Approved Standard Specification
		– Bursting strength	
9	الأغطية والعبايات الجراحية (Clean air suits)	– Resistance to microbial penetration – cleanliness microbial/bioburden	EN 13795 – part 2
10	المعقمات والمطهرات الكيميائية chemical disinfectants and antiseptics	– Disinfecting efficiency (antimicrobial efficacy) – Evaluation of fungicidal or yeasticidal activity – Evaluation of virucidal activity in the medical area	UAE.S EN 1276 EN 1650 EN 14476:2013+A2
11	طلاء الأسطح المضادة للفيروسات Anti-microbial surface coating	– Measuring antiviral activity	ISO 21702

Notes:

- Detergent products, cosmetics, and personal care products shall not be subject to the technical requirements stipulated in item (10) of the Annex.
- The technical requirement specified in item (10) of the Annex concerning the Evaluation of virucidal activity shall be met only if antiviral claims are made for the product.
- Reference is made to other international standard specifications not referred to in the Annex for the purpose of ensuring that the product's compliance with the technical requirements set out therein after their acceptance and approval by the Ministry.